

airpointer[®] CO module

IR Absorption

Lambert-Beer's Law:


$$I = I_0 e^{-\alpha Lc}$$




$$c = \ln(I_0/I) (1/\alpha L)$$

- I: Intensity of the light transmitted (after absorption)
- I_0 : Intensity of the light emitted
- α : absorption coefficient of the gas (here: CO)
- L: absorption path length
- c: concentration of the absorbing gas (here: CO)

Flow Diagram




Flow Diagram





GFC wheel

Unfortunately, water and CO₂ also absorb light with wavelength around 4.7 μ m. The CO Module uses a method called Gas Filter Correlation (GFC) to overcome their interfering effect.




NB: Sapphire will pass the 4.7 μ m photons. Normal glass is opaque at 4.7 μ m.


GFC wheel




GFC wheel




GFC wheel




GFC wheel



GFC wheel




GFC wheel




A change in CO_2 or H_2O concentration will impact both M and R: the M/R ratio stays constant

→ M/R does not depend on H_2O and CO_2 concentration


GFC wheel



GFC wheel



GFC wheel




A change in CO concentration will impact M, but not R: the M/R ratio changes


→ M/R depends on CO concentration

CO bench


Increased path length = increased sensitivity




CO bench



CO bench




Flow Diagram



Parameters

Parameter		Value		Unit		Status: BS-FS-SS							
CO		1.151		ppm		0 0 0							
CO_all	1.151	ppm	CO_raw	1.250	ppm	COStdDev	0.0442	ppm	CO_Avg (300 sec)	1.190	ppm	CalRatio (300 sec)	1.12823
COMeas	2073.2	mV	CORatio	1.1282	-								
CORef	1837.1	mV											
CO_AGC	5.04	V	Setpoint AGC	5.0 (+/- 0.2)	V								
CO_Speed	1990	rpm	PreAmpCO	40.4	%								
COIRSourceVoltage	18.7	V	COPowerToSpeed	65.2	%								
PressCO	879.0	mbar	FlowCO	502.5	ml/min								
BenchT	50.1	°C	PowerToCOBench	19.7	%								
COScrubberTemp	69.9	°C	PowerToCOScrubber	22.1	%								
CO_cylinder	1.0	bar											
CO Time Constant nr values to TC:				1200		StdDev last 10 samples:				0.001			
CO Slope:				1.098		CO Offset:				-0.060498			

Parameters




CO 1.306 ppm



Note: '-9999' is displayed for a missing value.

Span check



CO 1.306 ppm



Note: '-9999' is displayed for a missing value.

Calibration

- Using external zero air and external SO₂ cylinder



Preventive maintenance

- Change DFU filter (once a year)
- Change IR source (every 2 years)
- Never touch the mirrors inside the bench!

Full schedule available here:

<https://www.airpointer.tech/maintenance-schedule/>

Troubleshooting CO



Thank you for your attention!