

Base Unit - Replacing the sample fan (2D and 4D Airpointers)

- 1- Tools you will need
 - Torx 25 screwdriver
 - Phillips medium size screwdriver


- 2- Unscrew the 6 screws holding the fan exhaust grilles




- 3- Unscrew the 2 screws at the bottom of the Air Conditioner



- 4- Disconnect the microfit connector of the fan



- 5- Take out the fan and the fan holder



6- Change the fan (inbus 2.5)



6- Repeat steps 5 to 2 in the reverse order